
schülke


Gebrauchsanleitung

Point of Care (PoC) Halter 500 ml

1. Montage an horizontalen Rohrrahmen

Um den Halter an einem horizontalen Rohrrahmen zu befestigen, nutzen Sie den vertikalen Befestigungsriemen. Ein Anwenderhinweis auf horizontale Befestigung ist auf dem Riemen vorhanden. Führen Sie den Riemen von oben um die Rohrstange herum und durch die Umlenk-Öse (**Abb. I**). Drücken Sie den oberen Knopf durch die passende Öse, um den Riemen zu schließen (**Abb. II**). Der Riemen sollte so eng wie möglich gezogen werden, damit die maximale Haltekraft vorliegt.






Als nächstes bringen Sie den horizontalen Riemen in seine „Parkposition“: Führen Sie den Riemen hinter den Halter Richtung Front, dann wieder nach hinten durch die Umlenk-Öse (**Abb. III**) und verstauen Sie ihn an der Rückseite des Halters, damit er bei der Befestigung nicht stört (**Abb. IV**). Alternativ kann der ungenutzte Riemen abgeschnitten werden. Schließlich setzen Sie die Flasche ein.






2. Montage an vertikalen Rohrrahmen

Um den Halter an einem vertikalen Rohrahmen zu befestigen, nutzen Sie den horizontalen Befestigungsriemen. Ein Anwenderhinweis auf vertikale Befestigung ist auf dem Riemen vorhanden. Führen Sie den horizontalen Riemen um die Rohrstange herum (**Abb. V**) und drücken Sie den unteren Knopf durch die passende Öse, um den Riemen zu schließen (**Abb. VI**). Der Riemen sollte so eng wie möglich gezogen werden, damit die maximale Haltekraft vorliegt.





Als nächstes bringen Sie den vertikalen Riemen in seine „Parkposition“: Legen Sie ihn nach vorne und fixieren Sie ihn durch die dafür vorgesehene Öse am Kopf des Halters (**Abb. VII**). Alternativ kann der ungenutzte Riemen abgeschnitten werden. Schließlich setzen Sie die Flasche ein (**Abb. VIII**).





3. Flaschengröße

Der PoC Halter ist zur Aufnahme von 500 ml schülke Flaschen geeignet. Flaschen anderer Hersteller sind nicht auf die Kompatibilität mit dem Halter geprüft.

4. Wartung

Der PoC Halter ist wartungsfrei.

5. Pflege / Desinfektion

Den PoC Halter regelmäßig mit einem weichen Desinfektionstuch (z.B. mikrozid® universal wipes *plastic free wipes*) abwischen.

Der PoC Halter kann thermisch bis 60° C aufbereitet werden (z. B. mit thermosept® X·tra enzymatic cleaner).

Sicherheitshinweise: Nur an runden Rohren oder Stangen mit einem Durchmesser von 18 bis 40 mm befestigen. Nicht an Vierkantprofilen oder an Rohren mit größerem Durchmesser verwenden. Nur für 500 ml-Pumpflaschen verwenden. Keine größeren oder schwereren Behälter einsetzen. Nach der Montage den festen Sitz von Halter und Flasche prüfen. Halter regelmäßig auf Risse, Verfärbungen und Verformungen prüfen und bei Auffälligkeiten austauschen. Von Kindern fernhalten. Nicht zweckfremd als Spann- oder Gummiband verwenden. Entsorgung als sortenreiner Kunststoffabfall entsprechend den örtlichen und den betrieblichen Vorgaben. Kontaminierte Produkte sind nach den Hygiene- und Abfallvorgaben der Einrichtung zu entsorgen.

Einführer: HINAMIC GmbH · Rex-Brauns-Str. 11 · 31139 Hildesheim, Deutschland · mail@hinamic.de
Schülke & Mayr GmbH · Robert-Koch-Str. 2 · 22851 Norderstedt, Deutschland · info@schuelke.com

www.schuelke.com

901423 | III | 11.2027 | C | westwerk


schülke

Manual

Point of Care (PoC) Holder 500 ml

1. Mounting on horizontal tubular frames

To attach the holder to a horizontal tubular frame, use the vertical fastening strap. The strap is provided with a user indication for horizontal fastening. Guide the strap from above around the tube bar and through the deflection eyelet (**Fig. I**). Push the upper button through the corresponding eyelet to close the strap (**Fig. II**). The strap should be pulled as tight as possible for maximum holding force.






Then bring the horizontal strap into its 'parking position': guide the strap behind the holder towards the front, then backwards through the deflection eyelet (**Fig. III**) and stow it at the rear side of the holder so that it does not interfere with the attachment (**Fig. IV**). Alternatively, the unused strap can be cut off. You can then insert the bottle.






2. Mounting on vertical tubular frames

To attach the holder to a vertical tubular frame, use the horizontal fastening strap. The strap is provided with a user indication for vertical fastening. Guide the horizontal strap around the tube bar (**Fig. V**) and press the lower button through the corresponding eyelet to close the strap (**Fig. VI**). The strap should be pulled as tight as possible for maximum holding force.





Then bring the vertical strap into its 'parking position': place it forwards and secure it through the eyelet provided on the head of the holder (**Fig. VII**). Alternatively, the unused strap can be cut off. You can then insert the bottle (**Fig. VIII**).





3. Bottle size

The PoC holder is suitable for holding 500 ml schülke bottles. Bottles from other manufacturers have not been tested for compatibility with the holder.

4. Maintenance

The PoC holder requires no maintenance.

5. Care

Wipe down the PoC holder regularly with a soft disinfection wipe (e.g. mikrozid® universal wipes *plastic free wipes*).

The PoC holder can be thermally reprocessed up to 60° C (e.g. with thermosept® X·tra enzymatic cleaner).

Safety instructions: Attach only to round tubes or rods with a diameter of 18 to 40 mm. Do not use on square profiles or on tubes with a larger diameter. Use with 500 ml pump bottles only. Do not insert larger or heavier containers. After installation, ensure that both the holder and the bottle are securely attached. Regularly inspect the holder for cracks, discoloration, or deformation. Replace the holder if any damage is detected. Keep out of reach of children. Do not use for purposes other than intended use, such as a tension strap or rubber band. Dispose the holder as plastic waste in accordance with local regulations and facility procedures. Contaminated products must be disposed of in accordance with the facility's hygiene and waste management requirements.

Importer: HINAMIC GmbH · Rex-Brauns-Str. 11 · 31139 Hildesheim, Germany · mail@hinamic.de
Schülke & Mayr GmbH · Robert-Koch-Str. 2 · 22851 Norderstedt, Germany · info@schuelke.com

www.schuelke.com

901423 | III | 08.2026 | C | westwerk